

Infrastructure Commission (INFCOM)

Standing Committee on Measurements, Instrumentation and Traceability (SC-MINT)

Expert Team on Quality, Traceability and Calibration (ET-QTC)

Calibration of Wind Instruments

Part-5: Calibration Examples

Holger Doerschel (DWD)
Zafer Turgay Dag (TSMS)



WORLD
METEOROLOGICAL
ORGANIZATION

Content

Part 1:

- Introduction to this topic and historical information
- Concepts and definitions

Part 2:

- Methods of measurement

Part 3:

- Requirements for the laboratory setup #1

Part 4:

- Requirements for the laboratory setup #2

Part 5:

- **General procedure**
- Calibration example #1
- Calibration example #2
- Calibration example #3
- References and links

Calibration of anemometers

General procedure cup anemometer and ultrasonic anemometer

- calibration must be carried out over a range of values
- calibration must be carried out for both rising and falling wind speeds if hysteresis is to be expected
- the complete anemometer shall be placed in the wind tunnel with the measuring section
- environmental condition of the laboratory should be adjusted according to environmental criteria
- the device under test should last at least 24 hours for acclimatization

Calibration of anemometers

General procedure cup anemometer and ultrasonic anemometer

- device under test is installed at the test point that is decided before
- set up the Wind tunnel 10% U_{max} and wait approximately 5 min
- set up the Wind tunnel U_{max} (or max speed value according to capability of the wind tunnel) and wait approximately 5 min
- set up the Wind tunnel to different values (calibration points) and record the values from the reference and device under test
- evaluate the measuring uncertainty of the calibration according to their uncertainty parameters at the each calibration points separately

Calibration of anemometers

Standard	Calibration Test Speeds		
ISO 17713-1:2007 Meteorology – Wind measurements – Part 1: Wind Tunnel test methods for rotating anemometer performance (BS ISO 17713- 1:2007)	Test speeds cover the range of intended use for the anemometer and the working range of the wind tunnel, which includes five points in the non-linear range of the anemometer (near threshold, U_0). Typical test speeds are:		
	2 times U_0	0.1 times $U_{\max}$	0.6 times $U_{\max}$
	3 times U_0	0.2 times $U_{\max}$	0.7 times $U_{\max}$
	4 times U_0	0.3 times $U_{\max}$	0.8 times $U_{\max}$
	5 times U_0	0.4 times $U_{\max}$	0.9 times $U_{\max}$
	6 times U_0	0.5 times $U_{\max}$	$U_{\max}$

- the threshold value U_0 is the speed at which the cup anemometer begins to rotate very slowly
- $U_{\max}$ is the maximum speed for which the cup anemometer is designed

Calibration of anemometers

calibration points on cup anemometer:

Sensor specifications;

Measuring range: 0.4 – 75 m/s Threshold value : 0.5 m/s

Calibration points according to ISO 17713 standard :

2 times U_0	→	1.0 m/s
3 times U_0	→	1.5 m/s
4 times U_0	→	2.0 m/s
5 times U_0	→	2.5 m/s
6 times U_0	→	3.0 m/s
0.1 times $U_{\max}$	→	7.5 m/s
0.2 times $U_{\max}$	→	15.0 m/s
0.3 times $U_{\max}$	→	22.5 m/s
0.4 times $U_{\max}$	→	30.0 m/s
0.5 times $U_{\max}$	→	37.5 m/s
0.6 times $U_{\max}$	→	45.0 m/s
0.7 times $U_{\max}$	→	52.5 m/s
0.8 times $U_{\max}$	→	60.0 m/s
0.9 times $U_{\max}$	→	67.5 m/s
$U_{\max}$	→	75.0 m/s

Calibration points should be selected according to the infrastructure capability of the laboratory, calibration scope of the laboratory and customer demands.

Calibration of anemometers

uncertainties type A:

- data from the reference for example LDA
- data from other sources like air density (temperature, pressure, humidity)
- data from the calibration object

reference	calibration object
"Row#";"Arrival Time [ms]";"Transit Time [us]";"Velocity U [m/s]";"Arrival Time{G2} [ms]";"Transit Time{G2} [us]";"Velocity V{G2} [m/s]";	05.10.2018;10:43:20.326;6.0Hz
1;19,803;20,9;0,52;24,681;218,5;0,15	05.10.2018;10:43:20.577;7.0Hz
2;19,838;16,9;0,56;197,870;218,5;-0,17	05.10.2018;10:43:20.827;7.0Hz
3;24,943;436,8;0,49;234,562;10,6;-0,17	05.10.2018;10:43:21.770;6.0Hz
4;89,658;6,8;0,48;234,677;218,5;0,46	05.10.2018;10:43:21.327;7.0Hz
5;107,822;400,6;0,51;234,862;16,0;-0,01	05.10.2018;10:43:21.577;7.0Hz
6;114,413;105,2;0,46;234,874;6,8;0,14	05.10.2018;10:43:21.827;6.0Hz
7;114,513;89,2;0,47;234,881;6,9;0,27	05.10.2018;10:43:22.770;7.0Hz
8;126,311;436,8;0,46;234,887;3,0;-0,16	05.10.2018;10:43:22.311;7.0Hz
	05.10.2018;10:43:22.577;6.0Hz
	05.10.2018;10:43:22.827;7.0Hz
	05.10.2018;10:43:23.770;7.0Hz

Calibration of anemometers

uncertainties type A:

Erfassung- und Auswertesoftware für mechanische Windsensorik: i>Thies Classic Schalensternanemometer																
Kalibrierschein										DWD-Nr.: 00182472		Kalibrierdatum: 05.10.2018				
Startzeit	Endzeit	nnn	LDA m/s	Stabw m/s	SU m/s	LDA_K m/s	nnn	KG Hz	Stabw Hz	SU Hz	FC m/s	Stabw m/s	SU m/s	Abw m/s	k=1 m/s	
10:43:49.020	10:44:18.987	9396	0,505	0,059	0,001	0,510	120	7,000	0,290	0,026	0,622	0,015	0,001	0,112	0,050	OK
10:47:22.002	10:47:51.998	13855	1,020	0,063	0,001	1,031	120	16,558	0,499	0,046	1,108	0,025	0,002	0,077	0,050	OK
10:52:16.003	10:52:45.997	11284	2,445	0,008	0,000	2,474	120	43,400	0,586	0,053	2,464	0,029	0,003	-0,011	0,050	OK
10:55:11.001	10:55:41.000	14435	4,927	0,014	0,000	4,987	120	91,850	0,682	0,062	4,882	0,034	0,003	-0,105	0,050	OK
10:57:06.001	10:57:35.996	11873	7,285	0,017	0,000	7,376	120	139,942	0,946	0,086	7,250	0,046	0,004	-0,125	0,050	OK
10:59:08.002	10:59:38.000	20413	9,744	0,026	0,000	9,866	120	193,208	0,969	0,088	9,840	0,047	0,004	-0,026	0,059	OK
11:02:05.001	11:02:34.995	12793	12,160	0,025	0,000	12,313	120	245,575	1,105	0,101	12,358	0,053	0,005	0,046	0,074	OK
11:04:07.003	11:04:36.976	12719	14,619	0,032	0,000	14,804	120	299,258	1,338	0,122	14,916	0,063	0,006	0,112	0,089	OK
11:06:28.001	11:06:58.000	11747	19,430	0,045	0,000	19,676	120	405,392	1,440	0,131	19,918	0,067	0,006	0,242	0,118	OK
11:08:32.006	11:09:01.996	11046	29,304	0,067	0,001	29,681	120	620,625	2,162	0,197	29,961	0,101	0,009	0,279	0,178	OK
11:10:18.004	11:10:47.997	16750	39,070	0,108	0,001	39,579	120	830,742	2,779	0,254	39,786	0,131	0,012	0,207	0,237	OK
11:15:40.005	11:16:09.998	18886	48,604	0,127	0,001	49,243	120	1041,050	2,756	0,252	49,744	0,131	0,012	0,501	0,295	OK
Die Abweichung zum Normal liegt unter Einbeziehung der erweiterten Messunsicherheit bei allen Messungen im zulässigen Toleranzbereich.																

Calibration of anemometers

uncertainties type B:

- uncertainty of reference sensor
- annual drift of reference sensor
- environmental effects (temperature, humidity, pressure)
- resolution of reference sensor
- resolution of test sensor
- repeatability of reference sensor
- repeatability of test sensor
- homogeneity of wind tunnel

Calibration of anemometers

uncertainties type B:

- hysteresis
- blockage effect
- turbulence effect
- swirl effect
- power requirements
- not well trained staff
- ..

Calibration of anemometers

uncertainties type B:

- stability of wind tunnel
- source of any other winds especially heating, blower, air conditioner etc...
- opened doors or windows
- carriage, transportation
- vibrations etc
-

Content

Part 1:

- Introduction to this topic and historical information
- Concepts and definitions

Part 2:

- Methods of measurement

Part 3:

- Requirements for the laboratory setup #1

Part 4:

- Requirements for the laboratory setup #2

Part 5:

- General procedure
- **Calibration example #1**
- Calibration example #2
- Calibration example #3
- References and links

Calibration of anemometers – example #1

cup anemometer calibration at $\sim 1 \text{ ms}^{-1}$:

- model:

$$(U)^2 = (U_{\text{ref_cal}})^2 + (U_{\text{drift}})^2 + (U_{\text{test resolution}})^2 + (U_{\text{repeatability}})^2 + (U_{\text{stability}})^2 + (U_{\text{correction}})^2$$

- reference sensor calibration uncertainty $\rightarrow$ type B normal
- drift of reference sensor $\rightarrow$ type B rectangular
- test resolution $\rightarrow$ type B rectangular
- repeatability $\rightarrow$ type B
- stability $\rightarrow$ type B
- correction $\rightarrow$ type B

Calibration of anemometers – example #1

cup anemometer calibration at $\sim 1 \text{ ms}^{-1}$:

- the recordings:

Reference Speed (m/s)	Reference Correction value (m/s)	Corrected Reference Speed (m/s)	Test Speed (m/s)	Test Area Correction Value (m/s)	Corrected Test Speed (m/s)
1,102	0,0224	1,080	1,089	-0,0117	1,077
1,102	0,0224	1,080	1,089	-0,0117	1,077
1,102	0,0224	1,080	1,089	-0,0117	1,077
1,101	0,0224	1,079	1,089	-0,0117	1,077
1,101	0,0224	1,079	1,096	-0,0117	1,084
1,102	0,0224	1,080	1,096	-0,0117	1,084
1,102	0,0224	1,080	1,089	-0,0117	1,077
1,101	0,0224	1,079	1,082	-0,0117	1,070
1,103	0,0224	1,081	1,089	-0,0117	1,077
1,103	0,0224	1,081	1,089	-0,0117	1,077
Arithmetic Mean Speed		1,079			1,078

Calibration of anemometers – example #1

cup anemometer calibration at $\sim 1 \text{ ms}^{-1}$:

- the result:

1,0 m/s CALIBRATION POINT UNCERTAINTY PARAMETERS	ESTIMATED VALUE	DISTRIBUTION	DENOMINATOR	STANDARD UNCERTAINTY	NUMERATOR	PARTIAL VARIANCE
Reference Uncertainty (m/s)	0,01413	normal	2	0,00707	1	4,99488E-05
Drift (m/s)	0,00676	rectangular	1,73	0,00390	1	1,52403E-05
Test resolution (m/s)	0,00100	rectangular	1,73	0,00029	0,5	8,33333E-08
Error Repeatability (m/s)	0,00127	normal	1	0,00127	1	1,61687E-06
System Stability (m/s)	0,00023	normal	1	0,00023	1	5,22512E-08
Test area correction (m/s)	0,00188	normal	1	0,00188	1	3,53440E-06
Total variance						7,04761E-05
Standard Uncertainty (m/s)						0,00840
Expanded Uncertainty (k=2) (m/s)						0,01679
Relative Uncertainty (%)						1,56

Content

Part 1:

- Introduction to this topic and historical information
- Concepts and definitions

Part 2:

- Methods of measurement

Part 3:

- Requirements for the laboratory setup #1

Part 4:

- Requirements for the laboratory setup #2

Part 5:

- General procedure
- Calibration example #1
- **Calibration example #2**
- Calibration example #3
- References and links

Calibration of anemometers – example #2

cup anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- model part 1

$$\Delta \text{windspeed} = \text{windspeed_DUT} - \text{windspeed_reference}$$

$$\begin{aligned} \text{windspeed_reference} = & \text{measurements_LDA} + \\ & \delta \text{calibration_LDA} + \delta \text{calibration_BSA} + \\ & \delta \text{drift_BSA} + \delta \text{resolution_LDA} + \\ & \delta \text{flow_correction} + \delta \text{drift_LDA} + \\ & \delta \text{alignment_LDA} + \delta \text{temperature_LDA} \end{aligned}$$

$$\begin{aligned} \text{windspeed_DUT} = & \text{measurements_DUT} + \\ & \delta \text{calibration_frequency_counter} + \\ & \delta \text{resolution_frequency_counter} \end{aligned}$$

Calibration of anemometers – example #2

cup anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- model part 2

$\Delta \text{windspeed}$

windspeed_DUT

windspeed_reference

result (windspeed_DUT - windspeed_reference)

DUT indication of wind speed

reference indication of wind speed

Calibration of anemometers – example #2

cup anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- uncertainties:

measurements_LDA

δ calibration_LDA

δ calibration_BSA

δ drift_BSA

δ resolution_LDA

δ flow_correction

δ drift_LDA

δ alignment_LDA

δ temperature_LDA

measurements of LDA

calibration of LDA

calibration of burst spectrum analyzer

drift of BSA

resolution LDA

flow correction due to the type of DUT

drift of LDA

alignment of LDA

temperature dependency of LDA

Calibration of anemometers – example #2

cup anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- uncertainties

measurements_DUT

δ calibration_frequency_counter

δ resolution_frequency_counter

measurements of DUT converted
from frequency to speed and
corrected

calibration of frequency counter
resolution of frequency counter

Calibration of anemometers – example #2

cup anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- result:

Quantity	Value	Stand. uncert.	Distribution	DoF	Sensit. coeff.	Uncert. contribution	Rel. contribution	Bar chart
windspeed_DUT	49,57892 m/s	0,02352 m/s						
windspeed_reference	49,7197 m/s	0,2398 m/s						
measurements_LDA	49,719728 m/s	0,001597 m/s	Normal	∞	-1,000	-0,001597 m/s	0,00 %	
δ calibration_LDA	0,0 m/s	0,05469 m/s	Normal	∞	-1,000	-0,05469 m/s	5,15 %	
δ calibrationg_BSA	0,0 m/s	0,01000 m/s	Normal	∞	-1,000	-0,01000 m/s	0,17 %	
δ drift_BSA	0,0 m/s	0,0005774 m/s	Rectangular	∞	-1,000	-0,0005774 m/s	0,00 %	
δ resolution_LDA	0,0 m/s	0,002887 m/s	Rectangular	∞	-1,000	-0,002887 m/s	0,01 %	
δ flow_correction	0,0 m/s	0,2009 m/s	Rectangular	∞	-1,000	-0,2009 m/s	69,57 %	
δ drift_LDA	0,0 m/s	0,05741 m/s	Rectangular	∞	-1,000	-0,05741 m/s	5,68 %	
δ alignment_LDA	0,0 m/s	0,08612 m/s	Rectangular	∞	-1,000	-0,08612 m/s	12,78 %	
δ temperature_LDA	0,0 m/s	0,05741 m/s	Rectangular	∞	-1,000	-0,05741 m/s	5,68 %	
measurements_DUT	49,57892 m/s	0,01285 m/s	Normal	∞	1,000	0,01284 m/s	0,28 %	
δ calibration_frequency_counter	0,0 m/s	0,01400 m/s	Normal	∞	1,000	0,01400 m/s	0,34 %	
δ resolution_frequency_counter	0,0 m/s	0,01386 m/s	Rectangular	∞	1,000	0,01386 m/s	0,33 %	
Δwindspeed	Value -0,1408 m/s	Comb. stand. uncertainty 0,2409 m/s	Effective degrees of freedom ∞				Sign. digits 4	
Result	Value -0,141 m/s	Expanded uncertainty $\pm 0,482 \text{ m/s}$	Coverage factor (Probability) 2,00 (95,45 %)				Distribution Normal	Sign. digits 3

$$\Delta \text{windspeed} = -0,141 \text{ ms}^{-1} \pm 0,482 \text{ ms}^{-1} (k = 2)$$

Content

Part 1:

- Introduction to this topic and historical information
- Concepts and definitions

Part 2:

- Methods of measurement

Part 3:

- Requirements for the laboratory setup #1

Part 4:

- Requirements for the laboratory setup #2

Part 5:

- General procedure
- Calibration example #1
- Calibration example #2
- **Calibration example #3**
- References and links

Calibration of anemometers – example #3

ultrasonic anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- model part 1

$$\Delta \text{windspeed} = \text{windspeed_DUT} - \text{windspeed_LDA}$$

$$\begin{aligned} \text{windspeed_LDA} = & \text{measurements_LDA} + \delta \text{calibration_LDA} + \delta \text{drift_LDA} + \\ & \delta \text{calibration_BSA} + \delta \text{drift_BSA} + \delta \text{resolution_LDA} + \\ & \delta \text{flow_correction} + \delta \text{alignment_LDA} + \\ & \delta \text{temperature_LDA} + \delta \text{reproduceability} \end{aligned}$$

$$\begin{aligned} \text{windspeed_DUT} = & \text{measurements_DUT} + \delta \text{resolution_DUT} + \\ & \delta \text{blocking_DUT} \end{aligned}$$

Calibration of anemometers – example #2

ultrasonic anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- model part 2

$\Delta \text{windspeed}$

windspeed_DUT

windspeed_reference

result (windspeed_DUT - windspeed_reference)

DUT indication of wind speed

reference indication of wind speed

Calibration of anemometers – example #2

ultrasonic anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- uncertainties:

measurements_DUT

δ resolution_DUT

δ blocking_DUT

measurements of DUT

resolution DUT

blocking effect / correction of LDA
measurements

Calibration of anemometers – example #2

ultrasonic anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- uncertainties:

measurements_LDA

δ calibration_LDA

δ calibration_BSA

δ drift_BSA

δ resolution_LDA

δ flow_correction

δ drift_LDA

δ alignment_LDA

δ temperature_LDA

δ reproduceability

measurements of LDA

calibration of LDA

calibration of burst spectrum analyzer

drift of BSA

resolution LDA

flow correction due to the type of DUT

drift of LDA

alignment of LDA

temperature dependency of LDA

reproduceability of anemometer calibration

Calibration of anemometers – example #3

ultrasonic anemometer calibration at $\sim 50 \text{ ms}^{-1}$:

- result:

Quantity	Value	Stand. uncert.	Distribution	DoF	Sensit. coeff.	Uncert. contribution	Rel. contribution	Bar chart
windspeed_DUT	48,9400 m/s	0,2173 m/s						
windspeed_LDA	48,4340 m/s	0,2562 m/s						
measurements_LDA	48,434000 m/s	0,004000 m/s	Normal	∞	-1,000	-0,004000 m/s	0,01 %	
δ calibration_LDA	0,0 m/s	0,04117 m/s	Normal	∞	-1,000	-0,04117 m/s	1,50 %	
δ drift_LDA	0,0 m/s	0,07746 m/s	Rectangular	∞	-1,000	-0,07746 m/s	5,31 %	
δ calibrationg_BSA	0,0 m/s	0,001732 m/s	Rectangular	∞	-1,000	-0,001732 m/s	0,00 %	
δ drift_BSA	0,0 m/s	0,0005774 m/s	Rectangular	∞	-1,000	-0,0005774 m/s	0,00 %	
δ resolution_LDA	0,0 m/s	0,002887 m/s	Rectangular	∞	-1,000	-0,002887 m/s	0,01 %	
δ flow_correction	0,0 m/s	0,1678 m/s	Rectangular	∞	-1,000	-0,1678 m/s	24,93 %	
δ alignment_LDA	0,0 m/s	0,01398 m/s	Rectangular	∞	-1,000	-0,01398 m/s	0,17 %	
δ temperature_LDA	0,0 m/s	0,02887 m/s	Rectangular	∞	-1,000	-0,02887 m/s	0,74 %	
δ reproduceability	0,0 m/s	0,1696 m/s	Rectangular	∞	-1,000	-0,1696 m/s	25,47 %	
measurements_DUT	48,94000 m/s	0,05700 m/s	Normal	∞	1,000	0,05700 m/s	2,88 %	
δ resolution_DUT	0,0 m/s	0,002887 m/s	Rectangular	∞	1,000	0,002887 m/s	0,01 %	
δ blocking_DUT	0,0 m/s	0,2097 m/s	Rectangular	∞	1,000	0,2097 m/s	38,96 %	
Value		Comb. stand. uncertainty		Effective degrees of freedom			Sign. digits	
Δ windspeed	0,5060 m/s	0,3360 m/s		∞			4	
Value		Expanded uncertainty		Coverage factor (Probability)			Sign. digits	
Result	0,506 m/s	$\pm 0,672 \text{ m/s}$		2,00 (95,45 %)			3	

$$\Delta \text{windspeed} = 0,506 \text{ ms}^{-1} \pm 0,672 \text{ ms}^{-1} (k=2)$$

Content

Part 1:

- Introduction to this topic and historical information
- Concepts and definitions

Part 2:

- Methods of measurement

Part 3:

- Requirements for the laboratory setup #1

Part 4:

- Requirements for the laboratory setup #2

Part 5:

- General procedure
- Calibration example #1
- Calibration example #3
- Calibration example #2
- **References and links**

References and links

- JCGM 200:2012
International Vocabulary of Metrology - Basic and general concepts and associated terms (2012)
<https://www.bipm.org/en/committees/jc/jcgm/publications>
- JCGM 100:2008(E) – in English
Evaluation of measurement data - **Guide to the expression of uncertainty in measurement** (2008)
<https://www.bipm.org/en/committees/jc/jcgm/publications>

References and links

- Guide to Instruments and Methods of Observation (WMO-No. 8)
<https://library.wmo.int/records/item/41650-guide-to-instruments-and-methods-of-observation?offset=4>
- VDI 3650 Part 3 wind measuring devices
<https://www.vdi.de/richtlinien/details/vdi-3650-einrichtungen-zur-sicherung-von-kranen-gegen-abtreiben-durch-wind>
- VDI 3786 Part 2 Wind
<https://www.vdi.de/en/home/vdi-standards/details/vdi-3786-blatt-2-environmental-meteorology-meteorological-measurements-wind>

References and links

- ISO 16622:2002 Meteorology — Sonic anemometers/thermometers — Acceptance test methods for mean wind measurements
<https://www.iso.org/standard/29291.html>
- ISO 17713-1:2007 Meteorology — Wind Measurements Part 1: Wind tunnel test methods for rotating anemometer performance
<https://www.iso.org/standard/31497.html>

Calibration of Wind Instruments
End of Part 5

Thank you.



WORLD
METEOROLOGICAL
ORGANIZATION

wmo.int